



Recommendations

The version of the transmitter with two remote diaphragm seals is recommended for the measurement of pressure differences when the hydrostatic pressure of the manometric fluid in the capillaries (which depends on the vertical spacing of the seals) is significantly less than the measuring range of the transmitter. The best metrological results are obtained when the applied capillaries are identical, as short as possible, and terminated with identical seals. At such a configuration additional temperature errors, related to the remote sealing, affect both of the measurement chambers of the differential pressure transmitter in the same way, and thus cancel each other out.



Example of measurement of the level in a pressure tank

Recommendations

The transmitter with a direct diaphragm seal (connected to the positive measurement chamber) and a remote diaphragm seal (connected to the negative chamber) is recommended for hydrostatic measurements of: levels, densities, phase boundaries and pressure differences (with differentiated height of pulse source points*).

In such a configuration, at ambient temperature changes, two opposite phenomena appear concurrently.

Thermal expansion causes the change in the volume (and hence also the change in density) of the manometric fluid in the capillary, which results in a change of the hydrostatic pressure related to the vertical spacing of the seals. This phenomenon is counteracted by the elastic reaction of the diaphragm of the upper diaphragm seal, which is displaced by the change in volume of manometric fluid. Based on tests and experiments, the Aplisens transmitters are provided with carefully selected seal diaphragms, which guarantee compensation of the errors resulted from the ambient temperature changes.

The best metrological results are obtained using assembly, which include DN 80, DN 100, A 109 and S-Comp diaphragm seals or S-Mazut, S-DIN and S-Clamp diaphragm seals with a diameter of at least 65 mm, where the length of the capillary is $(1...1.3) \times$ (vertical spacing of seals). It is recommended using identical diaphragm seals at the both upper and lower connection points.

* The difference in height of pulse source points, at which the hydrostatic pressure of the manometric fluid is comparable to or greater than the range of the transmitter.



Note: The appropriate configuration of the complete set of pressure transmitter, diaphragm seals and capillaries, as well as the proper selection of manometric fluid, depends on several factors, including the physical and chemical properties, temperature range of the medium, the vertical spacing of the diaphragm seals, the measuring range, static pressure range, range of ambient temperatures and the technical specifications for mechanical connection of the diaphragm seals to the pressure devices.



Application and construction

The differential pressure transmitter is applicable to the measurement of pressure differences of: gases, vapours and liquids in cases where it is necessary to use seals and the pressure pulse source points may be several metres apart. Typical applications include the hydrostatic measurement of: levels in closed tanks, densities and phase boundaries, and the measurement of a filter loss, pressure differences between media in pasteurisers etc. The available range of the diaphragm seals allows measurement at great majority of media. The active element is a piezoresistant silicon sensor separated from the medium by a distance sealing system. The special design of the measuring unit means that it can withstand pressure surges and overloads of up to 40 bar. The electronic circuits are enclosed in a casing with a degree of protection IP 65 or IP66.

Configuration

The settings of the following metrological parameters can be changed:

- the units of pressure in which the range is configured,
- + start and end points of the range, time constant,
- inverted characteristic (output signal 20 ÷ 4 mA).

Communication

The transmitter is configured and calibrated using a KAP-03 communicator, some other communicators (HART) or a PC using an HART/USB converter and Aplisens RAPORT 2 configuration software.

The data interchange with the transmitter enables the users the transmitter identification, as well as reading of the currently measured differential pressure value, output current and percent of range width.

Measuring ranges

Nominal	Minimum set range	Vertical spacing	Maximum set range width,	Static
measuring range		of diaphragm	considering the actual vertical spacing	pressure limit
(FSO)		seals	of the diaphragm seals (m)	
-160160 mbar	0,1 m H ₂ O	≤ 1,7 m	[1,6 + (vertical spacing of seals × 0,94)] m H ₂ O	40 bar
-0,50.5 bar	0,5 m H ₂ O	≤ 6 m	[5 + (vertical spacing of seals × 1,04)] m H ₂ O	40 bar
-1,62 bar	1,5 m H ₂ O	≤15 m	[20 + (vertical spacing of seals × 1,04)] m H ₂ O	40 bar
-1,616 bar	1 bar	≤15 m	16 bar	40 bar

CAUTION: The maximum vertical diaphragm seal spacing shown in the table applies to level measurement, ensuring that it is possible to set the zero point of the transmitter when the tank is empty. For measurements of density or phase boundaries (in the sugar, chemical or refinery industries) the vertical spacing of the diaphragm seals can be larger.

Metrological parameters

 $\leq \pm 0.1\%$ (FSO)

Accuracy

The other parameters as given in the sheet for the smart differential pressure transmitter APR-2000ALW.

Sealing effect errors – as given in the relevant diaphragm seal sheet in chapter III (Diaphragm Seals), concerning the distance seal.

NOTE: The additional absolute zero error due to ambient temperature can be compensated by configuring

the transmitter, seals and capillaries in accordance with the recommendations on pages II/ 20 and II/ 21.

Electrical parameters

As given in the sheet for the APR-2000ALW differential pressure transmitter.

Operating conditions

Operating temperature range (ambient temperature) -25...85°C Exia, IS version: -25...80°C

Exd, XP version: -25...75°C

Medium temperature range – as given in the appropriate diaphragm seal sheet (remote seal)



Ordering procedure

Model	Code				Description						
APR-2000					Smart di	Smart differential pressure transmitter					
	/ALW			With dis	With display, output 4-20mA + Hart						
Versions	/ALW/Safety				With dis	play,	output 4-20mA + Hart				
versions					PN-EN 6	Functional Safety certificate according to PN-EN 61508:2010 parts 1 ÷ 7,					
					PN-EN 6	2061	2008 + PN-EN 62061:2008/A1:2013-06	+ PN-E	N 62061:2008/A2:2016-01		
	/SS	/SS		Stainles	s ste	el housing					
	/Exia				(Ex)	1	/2G Ex ia IIC T4/T5 Ga/Gb				
					IECEx	Ex	ia IIC T4/T5 Ga/Gb				
	/Exia (/Exia (Da)		 (C.)	II 1/2G Ex ia IIC T4/T5 Ga/Gb						
						IМ	1 Ex ia I Ma (version with SS housing	I)			
					1505	Ex ia IIC T4/T5 Ga/Gb					
					IECEX	IECEX EX IA IIIC 1105°C DA Ex ia LMa (version with SS housing)					
	/Exd					11 1	/2G Ex ia/db IIC T6/T5 Ga/Gb	<u>`</u>			
					(Ex)	II 1.	/2D Ex ia/tb IIIC T85°C/T100°C Da/D	b			
						IM	2 Exd ia I Mb (version with SS housi	ng)	Packing gland available on		
					IECEx	EX Fx	ia/db/liC/T8/T5/Ga/Gb ia/tb/IIIC/T85°C/T100°C/Da/Db		loquost		
						Ex	db ia I Mb (version with SS housing)	, J			
	/Exd (2	2G)			···· (Ex)	11 2	G Ex ia/db IIC T6/T5 Gb	_]		
0						11 2 Ev	D Ex Ia/tb IIIC 185°C/1100°C Db		Packing gland available on request		
Certificates, options "				IECEx	IECEx Ex ia/t IIIC T85°C/T100°C Db						
	/Exia(Da)/Exd		Dual cer	Dual certification Exia(Da) and Exd							
	/IS		··· FM	FM IS Class I, Div 1, Groups A, B, C, D T4							
			CSA (F	M)	IS Class II, Div 1, Groups E, F, G T	4					
					EM	EM Class I,Zone 0, AEx ia IIC T4					
						Zone 20, AEx ia IIIC T100°C					
	/XP		CSA (F	CSA (FM) Ex ia IIIC T100°C Da							
			FM		XP Class I, Div 1, Groups A, B, C, I	D T5					
			CSA (F	M)	DIP Class II, Div 1, Groups E, F, G	T5					
			EM		Class I, Zone 1, AEx db/ia IIC T5		Packing gland available on				
					FIVI		Zone 21, AEx tb/ia IIIC T100°C		request		
				CSA (F	M)	Ex db/ia IIC 15 Gb Ex tb/ia IIIC T100°C Db					
	/IS/XP	/IS/XP			Dual cer	Dual certification IS and XP					
	/SA /100 bar /160 bar		Surge ar	Surge arrester for Exia version							
			Static pr	Static pressure 100 bar							
*- more than one option			Static pr	Static pressure 160 bar							
is available	/IP67.	/IP67		Protectio	Protection class IP67						
	/Hart /				Commun	nicati	Dense		Min. and many a		
		/-160-160	nhar		-160±16	0 mb	range ar (-16÷16 kPa)	0.1 m			
Nominal measuring range	ie	/-0.5÷0.5 h	ar		-0.5÷0.5	-0.5+0.5 bar (-50+50 kPa)		0,5 ml	0.5 mH2O		
i torninai moacaniig rang		/-1.6+2 bar		1,6÷2 b	-1,6+2 bar (-160+200 kPa) 1,5 mH2O		120				
		/-1,6÷16 bar		1,6÷16	-1,6+16 bar (-160+1600 kPa) 1 bar						
Measuring set range /÷ [required units]			Calibrate	Calibrated range in relation to 4mA and 20mA output							
		/(+)		Direct di	Direct diaphragm seal or remote diaphragm seal mounted on the (+) side of the					
		K=		Capillan	transmitter- code as given in the relevant diaphragm seal sheet						
Process connections	/(-)		Remote	Remote diaphragm seal mounted on the (-) side of the transmitter – code as given in							
			the relev	the relevant diaphragm seal sheet							
		K	=		Capillary	/ leng	gth on (-) side of transmitter				
Electrical connection (without marking)			Packing	Packing gland M20x1,5 Thread 1/2"NPT Female							
Accessories			/05	/FI25	Mounting	ı/∠"N a.bre	r I remale				
Other specification				/	Descript	ion o	f required parameters				

Standard display configuration

	Std. version Exia(Da), IS Exd, XP Exia(Da)/Exd, IS/XP Safety						
	Std. version Exia, Exia(Da), IS Exd, XP Exia(Da)/Exd, IS/XP Sat						
Backlight on	•	•	•				
Backlight off				•	•		
Other configuration of display has to be marked upon placing order. User has no possibility of switching backlight on/off							

To simplify the mathematical operations we introduce the density coefficient of the medium X_{ρ} .

$$X\rho = \frac{\rho_{\text{medium}} [g/cm^3]}{\rho_{\text{water at 4°C}} [g/cm^3]}$$

Since the density of water at 4°C is 1 g/cm³, the **density coefficient** X_{ρ} **is numerically equal to the density of the medium expressed in g/cm³**. To determine the hydrostatic pressure of a column of liquid in mm H₂O, it is sufficient to multiply the height of the column h [mm] by the density coefficient of the liquid X_{ρ} . Since it is easy to determine the hydrostatic pressure in mm H₂O and the transmitter can be configured in those units, in the descriptions of measurement methods given below we will make use of pressures expressed in mm H₂O and the density coefficient X_{ρ} .

Configuration of the transmitter to measure the level of liquid in a tank

The measurement task:

PLISENS

To convert a variation in the level of a liquid with density ρ = 0.87 g/cm³ between 0 and h_{max} to a variation in the output signal from 4 to 20 mA.



- 1. Install the transmitter in its working position on an empty tank.
- 2. Make the electrical connections of the transmitter, providing for the ability to use HART communication.
- 3. Connect the KAP-03 communicator, identify the transmitter and select the "configuration" function.

- 4. On the configuration menu select the "Reranging" procedure.
- 5. On the "Reranging" menu:
 - a) change the units of measurement to mm H₂O at 4°C; b) enter the values for the start (Xń × h_{min} [mm]) and end (X ρ × h_{max} [mm]) of the measurement range,
 - namely 0 and (0.87 h_{max} [mm]) respectively;
 c) to compensate for the hydrostatic pressure of the manometric fluid, the start of the measurement range should be set using regulated pressure; when subject to the action of only the manometric fluid (empty tank) the transmitter will shift the start and end-points of the range, compensating for the value of that pressure.

When the transmitter has been configured in this way it is ready to be used to carry out the given measurement task.

If it is not possible to empty the tank to configure the transmitter, the hydrostatic pressure of the manometric fluid should be calculated by multiplying the vertical spacing of the diaphragm seals by the density coefficient of the oil in the capillaries. This pressure should be taken into account when entering the values for the start and end of the range:

Start [mm H₂O] = -H [mm] × X_{poil} End [mm H₂O] =

= h_{max} [mm] × $X \rho_{measured \ liquid}$ – H [mm] × $X \rho_{oil}$

 ρ_{oil} for DC-550 oil is equal to 1.068 g/cm³ ρ_{oil} for AK-20 oil is equal to 0.945 g/cm³

Configuration of the transmitter to measure density of liquids



The measurement task:

To convert a variation in liquid density from $\rho_{min} = 0.6 \text{ g/cm}^3$ to $\rho_{max} = 1.2 \text{ g/cm}^3$ to a variation in the output signal from 4 to 20 mA, with the vertical spacing of the diaphragm seals equal to H = 3000 mm. The sealing system is filled with DC-550 oil with density $\rho_{oil} = 1.068 \text{ g/cm}^3$.

- 1. Calculate the value of the start of the range as follows: $H_{[mm]}$ × (Xp_min - Xpoil) =
 - = 3000 × (0.6 1.068) = –1404 [mm H₂O]
- 2. Calculate the value of the end of the range as follows: $H_{[mm]} \times (X\rho_{max} - X\rho_{oil}) =$
- = 3000 × (1.2 1.068) = 396 [mm H₂O]
- 3. Set the zero point of the transmitter with the diaphragm seals positioned at the same level.
- 4. Install the transmitter in its working position.
- 5. Make the electrical connections to the transmitter, providing for the possibility of using HART communication.



- 6. Connect the KAP-03 communicator, identify the transmitter and select the "configuration" function.
- 7. On the configuration menu select "Reranging" procedure
- 8. On the "Reranging" menu:
- a) change the measurement units to mm H_2O at 4^\circC ;
- b) enter the calculated values for the start (-1404) and end (396) of the range.

When the transmitter has been configured in this way it is ready to be used to carry out the given measurement task.

Note: If it is possible to fill the space between the seals with a liquid whose density corresponds to the start of the measurement range, the start of the range of the transmitter can be set using regulated pressure.

Measurement of phase boundary

The height of the phase boundary of liquids of different densities is determined by measuring the average density of the medium between the seals.

Example:

Calculate the measurement range start and end points for an APR-2000/ALW transmitter configured to measure phase boundary height in the range 0-1000 mm between liquids of density $\rho_1 = 0.7$ g/cm³ and $\rho_2 = 1.0$ g/cm³, where the vertical spacing of the seals H = 1600 mm. The sealing system uses DC-550 oil with a density of 1.068 g/cm³.



To determine the start of the measurement range, calculate the pressure difference at the transmitter when the tank is filled with the lighter liquid only:

1600 [mm] × (0.7 – 1.068) = –588.8 [mm H₂O]

To determine the end-point of the range, add the increase in pressure resulting from the appearance of a 1 metre column of the heavier liquid:

Additional remarks

The settings of the transmitter can be adjusted with reference to laboratory results from density measurements carried out on samples of the liquid being measured. This is most often necessary when the measurement takes place in a pipeline segment where the flow velocity of the measured liquid reaches several m/s.

Increasing the vertical spacing of the diaphragm seals widens the range and often improves measurement accuracy.

In planning the spacing of the diaphragm seals, ensure that the pressure difference at the transmitter lies within the basic range.

The maximum vertical spacing of the diaphragm seals (H) depends on the transmitter's basic range and the boundary values for the density of the measured liquid (pmin; pmax).

If $\rho_{min} < \rho_{oil} < \rho_{max}$, the seal spacing H should satisfy the following conditions:

$$H \text{ [mm]} \leq \frac{\text{lower boundary of range [mm H_2O]}}{X\rho_{\text{min}} - X\rho_{\text{oil}}}$$

$$\text{H}\left[\text{mm}\right] \leq \frac{\text{upper boundary of range }\left[\text{mm }H_2\text{O}\right]}{X\rho_{\text{max}}-X\rho_{\text{oil}}}$$

Example:

F

Determine the maximum vertical spacing of the seals for the APR-2000ALW/-10...10 kPa transmitter when measuring the density of liquid between 0.6 and 1.2 g/cm³. The sealing system uses AK-20 silicone oil with a density of 0.945 g/cm³.

The lower boundary of the range of the transmitter is -10 kPa = -1020 mm H₂O

H [mm]
$$\leq \frac{-1020}{0.6 - 0.945}$$
 ⇒ H [mm] $\leq \frac{-1020}{-0.345}$ ⇒
H [mm] ≤ 2957

The upper boundary of the range of the transmitter is +10 kPa = 1020 mm H₂O

$$H [mm] \leq \frac{1020}{1.2 - 0.945} \quad \Rightarrow \quad H [mm] \leq \frac{1020}{0.255} \quad \Rightarrow \\ H [mm] \leq 4000$$

In the example, both conditions are satisfied when the spacing of the seals is not more than 2957 mm.